Noninvasive Hemoglobin for Blood Donation Centers

Quick, noninvasive total hemoglobin (SpHb®) spot-check measurements with the handheld Rad-67™ Pulse CO-Oximeter®



Benefits of Noninvasive Hemoglobin (SpHb)

- > Pain-free, noninvasive hemoglobin measurements for donor comfort
- Noninvasive technology does not introduce risk of exposure to bloodborne pathogens for donor or staff
- > Efficient and cost-effective solution with no manual calibration or waste generated



Next Generation SpHb Technology

The following table represents the accuracy of SpHb measurements obtained using Rad-67 and tHb measurements using an invasive point-of-care device, each compared to a laboratory reference device.

Device	Subjects	Samples	Std Dev	Bias	A _{RMS} ¹
SpHb vs. Laboratory Hematology Analyzer	319	660	1.0	0.1	1.0
Invasive Point-of-care Device vs. Laboratory Hematology Analyzer	283	283	1.1	-0.3	1.2



EASY-TO-USE DEVICE FOR EFFICIENCY AND OPTIMIZED WORKFLOW

Spot-check measurements including SpHb, pulse rate, SpO2, and other measurements obtained in just a few simple steps.



Intuitive touchscreen allows users to quickly navigate the user interface



Spot-check results displayed with **signal quality indicators**



Feedback screens provide alerts regarding signal quality and possible solutions



Review historical results on device and sort by customized donor identifier and date/time of measurement

HD Display

- Bright LCD, colour display
- Automatic low power mode to conserve power

Compact Design

- Minimizes storage needs
- Easily transportable

Auto-Brightness

 Ambient light sensor automatically adjusts screen brightness to optimize visibility

Rechargeable Battery

- Li-ion Battery
- Up to 6 hours battery life²
- 6 hours charging time



RAD-67 SPECIFICATIONS

ORDERING INFORMATION
Rad-67PN 9794
PHYSICAL CHARACTERISTICS
Weight
ENVIRONMENTAL
Operating Temperature. 0-35° C (32-95° F) Atmospheric Pressure 540-1,060 mBar Operating Humidity. 10-95%, non-condensing

COMPLIANCE	
Safety Standard(s)	. ANSI/AAMI ES 60601-1, CAN/CSA C22.2 No. 60601-1, IEC/EN 60601-1, 3rd Ed.
Pulse Oximeter Standard(s) IEC Standard(s)	
Type of Protection	Class II (Internally Powered)
Degree of Protection	Type BF, Defib Proof-Applied Part
Mode of Operation (per IEC 60601-1)	Continuous Operation
Enclosure Degree of Protection	IPX4

¹ Arms accuracy is a statistical calculation of the difference between device measurements and reference measurements. Approximately two-thirds of the device measurements fell within +/- Arms of the reference measurements in a controlled study. ²This represents approximate runtime at the lowest indicator brightness and wireless functionality powered off using a fully charged battery.

For professional use. See instructions for use for full prescribing information, including indications, contraindications, warnings, and precautions.

